





CONVEXITY
SCIENTIFIC

Flyp™ Nebulizer—Cleaning and Maintenance

To prevent buildup of residual medication, it's important to clean Flyp™ after each use. Users should consult the user's guide for full instructions as well as visit flypnebulizer.com for additional information.



Flyp Tip: If a user reports a decreased flow output or lengthening treatment times, it's usually an indication of the holes becoming clogged. Distilled white vinegar is an excellent organic solvent to provide a deep cleaning. Simply fill the medication reservoir with a bit of the vinegar and power the unit on to run for one cycle. We recommend doing so in a well ventilated area.

Warranty Information for Flyp™ Resellers

What's the warranty on Flyp Nebulizer?

- ✓ Flyp™ Nebulizer is warranted for a period of thirty-six months (3 years) from the date of purchase. This warranty protects against defects in manufacturing and all warranties are based on typical usage.

Does the warranty cover removeable components and accessories?

- ✓ The warranty does not include loss or damage to the removable components (mouthpiece or reservoir) or accessories. Replacement parts are available for sale through your distributor partner.

Should a reseller accept returns or exchanges directly?

- ✓ No. Please instruct the purchaser or owner of Flyp™ to contact Convexity Scientific Inc for all warranty claims:

By Phone: 844.FLYP.NEB (844.359.7632) and select option #2

OR

By Email: info@convexityscientific.com

This will allow the Flyp customer service team to assess the reported issue and offer helpful Flyp-Tips directly.